

Original Research Article

Comparative evaluation of primary stability of commercially available self- drilling orthodontic titanium mini implants inserted in polyurethane bone block at different insertion angles using pull out test - An *in vitro* study

AbdulRehman Irfan Foujdar^{1*}, Chaitali Purushottam Rekhate², Sameer Patil³, Sheetal Potnis³, Arun Mhaske¹, Sandeep Jethe¹

¹Dept. of Orthodontics & Dentofacial Orthopedics, D.Y. Patil Dental School, Pune, Maharashtra, India

²Dept. of Orthodontics & Dentofacial Orthopedics, Nanded Rural Dental College & Research Center, Nanded, Maharashtra, India

³Dept. of Orthodontics & Dentofacial Orthopedics, Sinhgad Dental College & Hospital, Pune, Maharashtra, India

Abstract

Background: orthodontic mini implants are widely used nowadays in orthodontic practices. It is made up of various materials, due to its advantage of greater anchorage control and better patient compliance they play a pivotal role in orthodontic treatments. Primary stability of these mini implants at different angulations are tested in this study.

Aim & Objective: Aim of the present study is to evaluate and compare the primary stability of self-drilling orthodontic titanium mini-implants (JJ Orthodontics) inserted in polyurethane bone blocks at different insertion angles using a pull-out test.

Materials and Methods: A total of 24 newly purchased (1.6 x 8 mm) orthodontic titanium mini-implants were obtained and inserted in polyurethane foam bone blocks simulating cortical and cancellous bone layers having density of 50 PCF and 40 PCF respectively with help of implant driver manually. Each specimen was tested for pull-out strength using the Universal Testing Machine with a loading speed of 1 mm/min.

Results: The mean distribution of Pull out Strength for JJ orthodontics inserted in polyurethane bone block at 90° and 60° was 98.50 ± 12.10 and 141.19 ± 12.92 respectively. A high statistically significant difference was observed in pull out strength at 60° and 90° angle (P<0.001).

Conclusion: As the result concludes the pull-out strength at 60° is greater than 90°, which states insertion angle plays an vital role in the primary stability of the mini implant.

Keywords: Mini-Implant, Insertion Angle, Pull-Out Test, Polyurethane Block, Orthodontics, Anchorage.

Received: 16-02-2025; **Accepted:** 16-09-2025; **Available Online:** 29-09-2025

This is an Open Access (OA) journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: reprint@ipinnovative.com

1. Introduction

In orthodontic treatment, achieving precise tooth repositioning relies on appropriately applied forces and innovative tools. Orthodontists utilize a variety of advanced devices, incorporating materials like elastics, springs, wires, and metals, to facilitate desired tooth movements efficiently. Through strategic integration of these components, tailored treatment plans are crafted to produce favorable outcomes, enhancing the effectiveness of orthodontic interventions.

Newton's third law states that every action has an equal and opposite response. When pressure is applied on groups of teeth, all of them should move to varying degrees.¹⁻⁴ The reactive pressures created by the appliance system may cause the teeth to move in an undesired fashion since the majority of orthodontic appliances are tooth-borne.

In orthodontics, Proffit describes anchoring as "resistance to undesirable tooth motions."⁵ Historical anchorage methods included intra-oral acrylic pads or extra-oral devices like headgear. Challenges arise when anchorage

*Corresponding author: AbdulRehman Irfan Foujdar
Email: iamdamad@gmail.com

units shift, necessitating precise anchoring to prevent undesired tooth movements. Absolute anchorage, achieved through mini implants, offers predictability, minimal invasiveness, and cost-effectiveness in orthodontic treatment.^{6,7} Various sizes of mini-implants are available in which larger mini-implants offer clinical benefits by spreading the force over wider bone areas, reducing bone stress.⁸

The efficacy of implant anchoring relies on primary stability upon insertion, influenced by factors such as quality of bone, design of implant, insertion technique, bone thickness, and angle of insertion. Studies indicate that inclined mini-implants improve contact with cortical bone, enhancing mechanical retention and stability.⁹

Mini-implants can achieve prompt loading with sufficient initial stability, which is vital for success. Factors influencing this stability include patient characteristics, implant design, and clinical technique.¹⁰ Insertion technique, including angle and predrilling, affects primary stability. It has been observed that mini-implants at 60–70 degrees to the bone surface show better stability than those at 90 degrees.^{11,12}

In orthodontics, the pull-out test assesses the retention properties of materials or devices, measuring the shear strength of mini-screws in bone specimens. Previous studies have examined pull-out loads to evaluate stability, considering factors like cortical thickness, healing times, and implant orientation in resisting failure.¹³

Previous research has explored factors such as length and diameter about the stability of mini implants, but there's been limited literature available on how insertion angles affects the stability. Therefore, this study aims to evaluate and compare the primary stability of orthodontic titanium mini-implants (JJ Orthodontics) inserted in polyurethane bone blocks at distinct insertion angle i.e 60° and 90°.

2. Materials and Methods

2.1. Inclusion criteria

1. Newly purchased mini-implants.
2. Mini-Implants with 8mm length and 1.6mm diameter.
3. Manufactured by JJ Orthodontics.
4. Made of titanium alloy.
5. Self-drilling mini-implants.

2.2. Exclusion criteria

1. Different dimensions or materials.
2. Not manufactured by the specified commercial houses.
3. Previously used mini-implants.
4. Self-tapping mini-implants.

2.3. Study instruments

1. Implant driver. (**Figure 1**)
2. New 1.6 x 8 mm self-drilling orthodontic titanium mini-implants from JJ Orthodontics. (**Figure 2,3**)
3. Synthetic Bone (Sawbones). (**Figure 4**)
4. Guide Block for 60° angle placement. (**Figure 5**)
5. Protractor.
6. Universal testing machine. (**Figure 6**)

2.4. Study setting and design

This in-vitro study was carried out in the Department of Orthodontics and Dentofacial Orthopedics. Sample size of 24 was obtained.¹⁰ A total of new 24 (1.6x8mm) orthodontic, self-drilling titanium mini-implants were purchased from JJ Orthodontics. Mini implants made up of different dimensions, materials, not manufactured from same company, previously used, self-tapping system was excluded from the study. Armamentarium such as implant driver, synthetic bone, and guide block for correct angulation was prepared and kept ready. These implants were conical formed, hexagonal design with small hole at the top for the passage of threads and ligature wires. Based on degree of angulation these implants were subdivided into Group 1A (60° angles) and Group 1B (90° angles) consisting of 12 each. Polyurethane foam bone blocks meeting ASTM F-1839-08 standards and simulating the cortical and cancellous bone layer of densities 50 PCF and 40 PCF. Bone block had two layers, 8.5 mm layer of cancellous bone and 1.5 mm of cortical bone. Both the group of twelve mini implants from JJ Orthodontics were inserted in the polyurethane bone block manually at 90° and 60° angle respectively. (**Figure 7,8**) Then the bone block was placed under the universal testing machine and was tied with the help of a metal clamp. Afterwards, the implant's head was tied to the universal testing machine's claw to assess the pull-out strength. The reading of the pull-out strength was measured in Newton for each specimen.

2.5. Statistical analysis

The data were analyzed using Microsoft Excel and IBM SPSS version 13. Descriptive statistics were applied to determine the mean pull-out strength. Unpaired t-test was used to compare pull-out strength among different implant houses and bone block densities. Independent t-tests were used for further comparisons. A 95% confidence interval and p-value <0.05 were considered statistically significant.

3. Result

The mean distribution of pull-out strength varied between different titanium mini-implants inserted in polyurethane bone blocks at different insertion angles. At a 90° angle, the pull-out strength was (98.50 ± 12.10). (**Table 1**) Conversely, at a 60° angle, the mean distribution of the pull-out strength was (141.19 ± 12.92). (**Table 2**)

There was a statistically significant difference in pull-out strength observed at 60° and 90° angles ($P < 0.001$) (**Table 3**).



Figure 1: Implant driver



Figure 2: Packaging of JJ orthodontic implant



Figure 3: JJ Orthodontic implant



Figure 4: Sawbones polyurethane foam (artificial bone)




Figure 5: Guide block for 60° angle placement



Figure 6: Universal testing machine



Figure 7: JJ Orthodontic implant places in 40 PCF artificial bone block at 90°




Figure 8: JJ Orthodontic implant places in 40 PCF artificial bone block at 60°

Table 1: Mean Distribution of Pull out Strength between Different titanium mini- implants from three different commercial houses inserted in polyurethane bone blocks at 90°

	N	Minimum	Maximum	Mean	Std. Deviation
90° JJ Orthodontics	12	77.20	112.60	98.5042	12.10667

Table 2: Mean Distribution of pull out strength between different titanium mini- implants from three different commercial houses inserted in polyurethane bone blocks at 60°

	N	Minimum	Maximum	Mean	Std. Deviation
60° JJ Orthodontics	12	125.00	165.50	141.1917	12.92408

Table 3: Comparison of Pull out Strength between JJ Orthodontics titanium mini- implants inserted at different angles

Group Statistics							
	Groups	N	Mean	Std. Deviation	Mean Difference	t	P Value
Pull-out strength	90° JJ Orthodontics	12	98.5042	12.10667	-42.68750	-8.350	.000
	60° JJ Orthodontics	12	141.1917	12.92408			

4. Discussion

In orthodontics, ensuring the stability of implants or mini-implants is paramount for successful treatment outcomes. Primary stability, characterized by the absence of motion in the bone bed immediately after insertion, is crucial for preventing gradual mobility and potential loss of the device. This stability hinges on various factors, including device attributes, insertion method, and bone quality at the receptor site. In clinical practice, the initial stability of mini-screws is also vital, especially considering the early load applied to them in many patients. Recommendations include replacing inadequately retained mini-implants with thicker devices or altering insertion points to ensure optimal mechanical retention.¹⁴ The insertion angle of implants in the bone is one of the most important factors controlling the primary stability of temporary skeletal anchorage devices.

Haneen et al.¹⁵ employed a pull-out test to assess the primary stability of mini-implants, a common method for evaluating embedding retention characteristics. This test measures the pullout force, indicating the shear strength of mini-screws inserted into bone specimens. Their study investigated the stability of mini-implants at different insertion angles in synthetic bone. The same method was used in this study, where researchers evaluated the pullout strength of mini implants at two insertion angles. To assess primary stability, they used artificial bone from Sawbones, chosen for its homogeneity to ensure standardized and reproducible testing, facilitating comparisons across a large sample size. Among various synthetic bone options, polyurethane foam, generated through a polymerization reaction of water and isocyanate, is extensively used in orthopedic experiments due to its closed-cell structure. Approved by the American Society for Testing and Materials (ASTM F-1839-08), polyurethane is recognized as a standard material for testing instruments and oral implants, facilitating comparative evaluations of bone screws. Its uniformity and consistent material properties make rigid polyurethane ideal for conducting relative tests of various medical devices and implants.¹⁶

Bone mineral density (BMD) values of the mandible range from 0.528 to 0.820 g/cm³, averaging 0.661 g/cm³, while maxillary BMD ranges from 0.31 to 0.55 g/cm³. Cortical bone thickness varies from 1.09 to 2.12 mm in the mandible and 1.59 to 3.03 mm in the maxilla. To simulate both cortical and cancellous bone, a bone block with a 1.5mm cortical layer (density: 0.80g/cm³ or 50 PCF) and an 8.5-mm cancellous layer (density: 0.64g/cm³ or 40 PCF) was used. Thus, customized Sawbones with a 1.5mm cortical layer (density: 50 PCF) and 40 PCF cancellous density were selected for the study.¹⁵

To the best of our knowledge, very few literature is available on the comparison of primary stability of different self-drilling orthodontic titanium mini at different insertion angle i.e. 90 degree and 60 degree in synthetic bone block

The present study showed that the pullout strength of group 1 (at 60°) was significantly higher when compared with the pullout strength of group 2 (at 90°) for JJ Orthodontics. Similarly, in a study by Araghbidikashani et al.,¹⁷ they evaluated the pull-out strength at different insertion angles, which is similar to this study, but they used a different brand of implant and also placed it in the sheep mandible, which is different from our study. The result of the study was statistically significant, and the pullout strength was higher than in this study.

In another study, Meira T.M. et al.⁹ evaluated the pull-out strength of mini-implants at different insertion angles (45°, 60° and 90°) with sizes of 1.6 mm in diameter and 7 mm in length inserted in polyurethane bone blocks. In contrast to our findings, they found that there is no statistically significant difference in the pull-out strength of mini-implants inserted at 60° and 90°, whereas in the current study, there was a statistically significant difference in the pull-out strength of mini-implants. The difference in the result might be due to variations in the sizes and brands of mini implants used in this study.

In a FEM study by Lee J et al.,¹⁸ a bone was constructed with a 1mm thickness of cortical bone and a 10mm thickness of cancellous bone. The combination-type scheme with

cylindrical and tapered parts served as the basis for the miniscrew design (Orlus, Ortholution, Seoul, Korea), and they were inserted at different angles in the bone. When pull-out strength was evaluated in the study, it showed a statistically significant difference in pull-out strength. In the current study, there was also a statistically significant difference in the pull-out strength at different insertion angles

Mohammed et al.¹⁵ in an in-vitro study evaluated the pull-out strength of three different commercially available mini-implants of two different lengths (8mm and 10 mm) and two different diameters (1.6mm and 1.8mm) placed at a 90° insertion angle in the artificial bone block. He found a statistically significant difference between the pull-out strength of 1.6 mm in diameter and 8 mm in length when placed at 90° which bears resemblance to the present study. There was a difference observed in the mean pullout strength in both studies as different commercial house implants were used.

The pull out strength at 60 was more than 90 for JJ Orthodontic mini-implants. The explanation for this outcome could be because of the varied mini-implant body design, pitch of the implants, distance between the head and thread.

5. Strengths of the Study

To our knowledge, this is the first study conducted using these particular bone densities and commercial houses of mini implants to compare the primary stability of mini implants at different insertion angles using pull-out strength.

As synthetic bone was used, the results can be generated for possible comparison, and variation in results due to bone density is avoided.

The head diameter and length of the mini implant are kept the same for all the commercial houses, so bias due to these differences can be avoided, though factors like thread design and shape of the mini implants need to be considered.

6. Limitations

It cannot be assured that the outcomes of the in-vitro study can be applied to the clinical situation as it was conducted in a laboratory setting, which is its fundamental shortcoming.

If the study is performed in vivo, age, sex, hormonal changes, bone matrix composition, certain disorders, and other variables all affect bone density; thus, they must be taken into account.

Since characteristics such as implant thread design and shape also affect primary stability, they must be taken into consideration.

7. Conclusion

In orthodontic therapy, achieving anteroposterior correction while maintaining anchorage is crucial. Mini implants serve

as anchorage entities, increasingly utilised to prevent space loss during treatment. However, mini-implant failure due to insufficient primary stability is common, emphasizing the importance of understanding the factors affecting it, such as implant design, insertion technique, and bone quality. This study evaluated the impact of insertion angle on primary stability. Results indicated that a 60° angle yielded greater pull-out strength than 90°, enhancing primary stability. This highlights the significance of the insertion angle and suggests further research is needed in this area.

8. Source of Funding

None.

9. Conflict of Interest

None.

References

1. Mehta F, Pathak K, Patel R, Bhattacharya A, Trivedi R, Sanjeliwala A. A comprehensive review: Micro-implants in orthodontics. *Int J Rec Sci Res*. 2018;9(12):29974–81.
2. Kuroda S, Sugawara Y, Deguchi T, Kyung HM, Takano-Yamamoto T. Clinical use of miniscrew implants as orthodontic anchorage: success rates and postoperative discomfort. *Am J Orthod Dentofacial Orthop*. 2007;131(1):9–15.
3. Huang LH, Shotwell JL, Wang HL. Dental implants for orthodontic anchorage. *Am J Orthod Dentofacial Orthop*. 2005;127(6):713–22.
4. Chen CH, Chang CS, Hsieh CH, Tseng YC, Shen YS, Huang IY, Yang CF, Chen CM. The use of microimplants in orthodontic anchorage. *J Oral Maxillofac Surg*. 2006;64(8):1209–13.
5. W R Proffit, H Fields, B Larson, D M Sarver, Contemporary Orthodontics - E-Book: Contemporary Orthodontics - E-Book. Elsevier Health Sciences, 2018.
6. Iturbe OI, Núñez EG, Gayosso CA, Ibarra JG. Resistance to traction forces in mini-implants used in orthodontics depending on the insertion angle. *Rev Mex Orthod*. 2014;2(3):187–91.
7. Baumgaertel S. Temporary skeletal anchorage devices: the case for miniscrews. *Am J Orthod Dentofacial Orthop*. 2014;145(5):558–64.
8. Tatli U, Alraawi M, Toroğlu MS. Effects of size and insertion angle of orthodontic mini-implants on skeletal anchorage. *Am J Orthod Dentofacial Orthop*. 2019;156(2):220–8.
9. Meira TM, Tanaka OM, Ronsani MM, Maruo IT, Guariza-Filho O, Camargo ES, Maruo H. Insertion torque, pull-out strength and cortical bone thickness in contact with orthodontic mini-implants at different insertion angles. *Eur J Orthod*. 2013;35(6):766–71.
10. Holm L, Cunningham SJ, Petrie A, Cousley RR. An in vitro study of factors affecting the primary stability of orthodontic mini-implants. *Angle Orthod*. 2012;82(6):1022–8.
11. Inaba M. Evaluation of primary stability of inclined orthodontic mini-implants. *J Oral Sci*. 2009;51(3):347–53.
12. Wilmes B, Su YY, Drescher D. Insertion angle impact on primary stability of orthodontic mini-implants. *Angle Orthod*. 2008;78(6):1065–70.
13. Miglioni M, Benedicenti S, Signori A, Drago S, Barberis F, Tournier H, Silvestrini-Biavati A. Miniscrew design and bone characteristics: an experimental study of primary stability. *Am J Orthod Dentofacial Orthop*. 2012;142(2):228–34.
14. Markezan M, Mattos CT, Sant'Anna EF, de Souza MM, Maia LC. Does cortical thickness influence the primary stability of miniscrews?: A systematic review and meta-analysis. *Angle Orthod*. 2014;84(6):1093–103.
15. Mohammad H, Al-Sheakli. Comparing the primary stability of three different orthodontic mini-implant with various dimensions on artificial bone. *Int J Med Res Health Sci*. 2018;7(1):128–34.

16. Comuzzi L, Tumedei M, Pontes AE, Piattelli A, Iezzi G. Primary Stability of Dental Implants in Low-Density (10 and 20 pcf) Polyurethane Foam Blocks: Conical vs Cylindrical Implants. *Int J Environ Res Public Health*. 2020;17(8):2617.
17. Araghibidikashani M, Golshah A, Nikkerdar N and Rezaei M. In-vitro impact of insertion angle on primary stability of miniscrews. *Am J Orthod Dentofacial Orthop*. 2016;150(3):436–43.
18. Lee J, Kim JY, Choi YJ, Kim KH, Chung CJ. Effects of placement angle and direction of orthopedic force application on the stability of orthodontic miniscrews. *Angle Orthod*. 2013;83(4):667–73.

Cite this article: Foujdar ARI, Rekhate CP, Patil S, Potnis S, Mhaske A, Jethi S. Comparative evaluation of primary stability of commercially available self- drilling orthodontic titanium mini implants inserted in polyurethane bone block at different insertion angles using pull out test - An *in vitro* study. *J Dent Spec*. 2025;13(2):279-284.